



INVITATION /REQUEST FOR SUBMISSION OF PRICE QUOTATION
Posted in PhilGEPS Webstie/Conspicuous Place at the Procureing Entity

The University of Science & Technology of Southern Philippines, through the Bids and Awards Committee (BAC) hereby invites interested parties to submit quotation for the procurement of the following item:

Reference No.	Qty.	Unit	Specification/Description	Budget for the Contract	Mode of Procurement
RFQ-NP-SVP-858-2022	Please see details below	Please see details below	Negotiated Procurement for the Technical Scientific Equipment	₱275,000.00	Small Value Procurement- Section 53.9, Rule XVI, Revised IRR of RA 9184)

Deadline for the submission of quotation is on November 16, 2022 at 12:00 noon at the Procurement Unit, 2nd Floor, USTP Gymnasium building. You may submit your quotation through fax number 856-4696 / 856-1739 or through personal delivery in a sealed envelope.

The University of Science & Technology of Southern Philippines reserves the right to accept or reject any or all bids / price quotations, to annul the process, and to reject all bids at any time prior to contract award, without thereby incurring any liability to the affected bidder or bidders.

ATTY. ERWIN B. BUCIO
BAC II Chairman

Date: November 10, 2022
 Quotation No.: RFQ-NP-SVP-858-2022

Name of Project: Procurement of the
 Technical Scientific Equipment

The Manager

Cagayan de Oro City

Please quote your lowest price on the item/s listed below, subject to the general Conditions stated below, stating the shortest time of delivery and submit your quotation duly signed by your representative not later than November 16, 2022 noon. Please enclosed in a sealed envelope.

ATTY. ERWIN B. BUCIO

BAC II Chairman

- NOTE:
1. DELIVERY PERIOD WITHIN 7 CALENDAR DAYS UPON RECEIPT OF PO.
 2. PRICE VALIDITY SHALL BE FOR A PERIOD OF 30 CALENDAR DAYS
 3. BIDDERS SHALL SUBMIT ORIGINAL BROCHURES SHOWING CERTIFICATIONS OF THE PRODUCT BEING OFFERED. (if applicable)
 4. THE USTP RESERVES THE RIGHT TO REJECT ANY OR ALL BIDS FOR ANY JUSTIFIABLE AND REASONABLE GROUND WHERE THE AWARD OF THE CONTRACT WILL NOT REDOUND TO THE BENEFIT OF THE GOP.
 5. REQUIREMENTS: 1) PHILGEPS REGISTRATION (PLATINUM MEMBERSHIP).

ITEM #	QTY	UNIT	DESCRIPTIONS AND SPECIFICATIONS	APPROVED UNIT COST (php)	UNIT PRICE
1	1	set	<p>TECHNICAL SCIENTIFIC EQUIPMENT (214)-LENSES</p> <ul style="list-style-type: none"> • Portable Air Quality Monitor (1set) -Portable, Indoor and Outdoor use -measures PM10, PM2.5 and CO2 levels in the air -PM10, PM2.5 accuracy: ±0.005mg/m3+15% -CO2 accuracy: <±20ppm+5%reading/ <±50ppm+5%reading -Measurement units Gas: ppm or mg/m3 Humidity: % Temperature °C or °F - Reading functions Instant, minimum, maximum, average - Rechargeable battery - Environmental operating conditions Temperature: -5°C to 45°C Humidity: 0 to 95% non-condensing -with Carry Case -with 1 extra battery -with Charger -with needed sensors for PM10, PM2.5, and CO2 • Environmental Meter (1 set) - Measures Humidity, Temperature, Air Velocity and Light. -Ergonomic pocket size housing with large dual LCD simultaneous display of Temperature and Air Velocity or Relative Humidity -Characters on display reverse direction depending on Hygro-Thermo-Anemometer or Light Mode -Data Hold to freeze the displayed value -Records Min/Max readings -Built-in low friction vane wheel improves accuracy of air velocity in ft/min, MPH, m/s, km/h, and Knots Built-in precision thin-film capacitance humidity sensor for fast response -Built in thermistor for ambient temperature measurements -Type K input measures Temperature to 2372°F (1300°C) using optional Type K probe -Utilizes precision photo diode and color correction filter -Cosine and color corrected light measurements -Defaults to last function when powered off -Auto power off 10 minutes after last key is pressed with disable feature -Low Battery and overrange indication -Complete with built-in Humidity sensor, Light sensor and Vane wheel, wrist strap, and 9V battery; Optional Type K thermocouple probe. • Sound Level Meter (1set) -Complete with Sound Meter, microphone wind screen, calibration screw driver, 9V battery, plus Sound Calibrator with 9V battery. Supplied in an attractive hard carrying case that provides protection and organization for your meter and accessories -High accuracy meets ANSI, and IEC Type 2 standards -High and Low measuring ranges (35 to 130dB) -Data Hold and Max Hold functions -Backlit display to view in dimly lit area -Calibrates and verifies sound level meter operation -Durable double molded housing -1kHz sine wave at 94dB/114dB is generated to an accuracy of 4% (frequency) and ±0.5dB -Total Harmonic Distortion (THD) <2% at 94dB <p>NOTE: PLEASE SEE ATTACHED SAMPLE PICTURES</p>	275,000.00	

Delivery Period:

Price Validity:

After having carefully read and accepted for General Conditions, I/We quote on the item at prices noted above.

Wabe-Lenses
RFQ-NP-SVP-858-2022
Resolution No. 1071 s. 2022
PR No. 202205-747

Tel. #./address/e-mail add

Date

C.M. Recto Avenue, Lapanan, Cagayan de Oro City 9000 Philippines, Tel. Nos. +63 (88) 856 1738; Telefax +63 (88) 856 4696 | www.ustp.edu.ph